



Society of Gastroenterology Nurses and Associates, Inc.

Facility CRE Risk Assessment Tool

The purpose of this tool is to provide a framework for assessing a facility's risk specific to CRE transmission related to ERCP procedures. It is not intended to be an all-inclusive list.

The response of **NO** to any of the items listed indicates further action(s) on the part of the facility.

Facility CRE Assessment		
Competency and Education	YES	NO
Do you have dedicated staff for endoscope reprocessing?		
Do you have a reprocessing protocol/policy at your facility?		
Are reprocessing and manufacturer guidelines readily available to endoscopy staff?		
Has every staff member had an orientation to endoscope reprocessing?		
Does your staff complete annual competencies for endoscope reprocessing e.g. ERCP scope?		
Do you have a process for reevaluating and updating training when there are any changes in reprocessing or manufacture's guidelines?		
Endoscopy / Facilities	YES	NO
Is the risk of CRE infection known and monitored at my institution?		
Do you have documentation to support tracking of endoscope use and reprocessing?		
Do you have an outbreak notification process in place?		
Do you have identified clean and dirty reprocessing areas?		
Do you have a culture/surveillance protocol for your unit?		
Do you have a dedicated staff member assigned to infection prevention?		
Do you perform an ongoing staffing needs assessment of reprocessing personnel related to procedure volume and type of scopes utilized?		

Disclaimer

The Society of Gastroenterology Nurses and Associates, Inc. (SGNA) assumes no responsibility for the practices or recommendations of any member or other practitioner, or for the policies and procedures of any practice setting. Nurses and associates function within the limitations of licensure, state nurse practice act, and/or institutional policy.